

DECLARATION OF CONFORMITY

Manufacturer's Name: Maxtec
Address: 2305 South 1070 West
Salt Lake City, Utah 84119 USA

European Representative: QNET BV
Kantstraat 19
NL-5076 NP Haaren
The Netherlands

Product: Oxygen Sensor

Model(s): MAX-1, MAX-11, MAX-11i, MAX-12, MAX-12A, MAX-12F, MAX-13,
MAX-13F, MAX-14, MAX-16, MAX-17, MAX-19, MAX-22cc &
MAX-250ESF

Classification & GMDN: IIa Sensor, Oxygen - 13538

Classification criteria: Clause 1.2 Rule 2 of Annex IX of MDD

We herewith declare that the above mentioned products meet the provisions of the following EC Council Directives and Standards. All supporting documents are retained under the premises of the manufacturer and the European Representative.


Directives: General application directives: Medical Device Directive, COUNCIL DIRECTIVE 93/42/EEC of 14 June 1993 per Annex II as amended by 2007/47/EC of 5 September 2007

Notified Body: TÜV SÜD Product Service
RIDLERSTRASSE 65, D-80339 MUNICH, Germany
Number 0123

EC Certificate No.: G1 16 10 45041 020

Date CE mark was affixed: 15 April 2001

This declaration is considered valid from December 19, 2016 to December 18, 2021.

Signature:  Date: 23 Jan 2017

Name: Lillian Erickson

Position: Director of Regulatory and Quality

Applied Standards

The referenced list of harmonized standards for which documented evidence of compliance can be provided includes:

EN 1441:2008
EN ISO 14971:2012
EN 60601-1:2006